

TECHNICAL SERVICE BULLETIN

#08-001



Date Issued: February 05, 2008

Product:

VALIDATE[®] Product	REF	LOT
Chem 6bc	106bc	6AA19007

Subject: Lot 6AA19007 Expiration Date reduced to 2008-03-31

Explanation:

Ongoing Quality Assurance testing has led us to shorten the expiration date on Chem 6bc Lot number 6AA19007. **The new Expiration date for Chem 6bc Lot# 6AA19007 is 2008-03-31. No other lots of Chem 6bc are affected by this change.**

Necessary Actions:

- Complete and return the attached fax back form so that we know this information was read and understood.
- Please share this information with the appropriate members of your department.
- Contact our Customer Support Department if you are unable to complete your second round of calibration verification / linearity testing due to this change.

We apologize for any inconvenience this caused your laboratory and look forward to your continued support of Maine Standards Company products.

Contact Information:

Customer Support Department
Maine Standards Company
765 Roosevelt Trail
Windham, ME 04062
Phone: 1-800-377-9684
Fax: 1-207-892-2266
E-mail: Info@mainestandards.com

February 2008

FAX BACK RESPONSE FORM

Please return this form by fax or mail to:
Maine Standards Company
Attn: Technical Support
765 Roosevelt Trail
Windham, ME 04062

Fax number: (207) 892-2266

RE: Technical Service Bulletin TSB #2008-001

VALIDATE[®] Product	REF	LOT
Chem 6bc	106bc	6AA19007

Complete the appropriate section(s) below:

I have received and understand TSB 2008-001 regarding the reduced expiration dating for Chem 6bc 6AA19007 and have updated my records appropriately.

or

I do not use / have Chem 6bc Lot# 6AA19007.

Your Facility: _____

Your Name (Print): _____

Your Job Title: _____

Your Telephone Number: _____

Your Signature: _____

Date: _____