

TECHNICAL SERVICE BULLETIN

#08-005



Date Issued: May 12, 2008

Product:

VALIDATE[®] Product	REF
CM 1ab	401ab

Subject: Addition of Myoglobin (MYO) claim to VALIDATE[®] CM 1ab test sets

Explanation: The VALIDATE[®] CM 1ab calibration verification / linearity product now can be used to test the reportable range of MYO assays.

The improvements will be implemented with Kit Lot # 40AA09208 Exp 2009-03-01 and newer.

Contact Information:

Customer Support Department
Maine Standards Company
765 Roosevelt Trail
Windham, ME 04062
Phone: 1-800-377-9684
Fax: 1-207-892-2266
E-mail: Info@mainestandards.com